

**ENGLISH****S-2238™**

For Laboratory Use Only

S-2238™

S-2238 is a chromogenic substrate for thrombin.

COMPOSITION

Each vial contains chromogenic substrate S-2238 25 mg and mannitol 120 mg as a bulking agent.

CHEMISTRY

Chemical name: H-D-Phenylalanyl-L-pipecoyl-L-arginine-p-nitroaniline dihydrochloride.

Formula: H-D-Phe-Pip-Arg-pNA · 2 HCl

Mol. wt.: 625.6

$\epsilon_{316\text{ nm}}$: $1.27 \cdot 10^4 \text{ mol}^{-1} \cdot \text{L} \cdot \text{cm}^{-1}$

Solubility: > 10 mmol/L in H₂O

Stability: Substance: Stable until expiry date if stored at 2-8°C. Avoid exposure to light. The substance is hygroscopic and should be stored dry. Solution: 1 mmol/L in H₂O is stable for more than 6 months at 2-8°C. Contamination by microorganisms may cause hydrolysis.

Suitable stock solution:

1-2 mmol/L in H₂O.**PRINCIPLE**

H-D-Phe-Pip-Arg-pNA $\xrightarrow{\text{Enzyme}}$ H-D-Phe-Pip-Arg-OH+pNA
The method for the determination of activity is based on the difference in absorbance (optical density) between the pNA formed and the original substrate. The rate of pNA formation, i.e. the increase in absorbance per second at 405 nm, is proportional to the enzymatic activity and is conveniently determined with a photometer.

KINETIC DATA

Human thrombin: $K_m = 0.7 \cdot 10^{-5} \text{ mol/L}$
 $V = 1.7 \cdot 10^{-7} \text{ mol/min} \cdot \text{NIH-U}$

Bovine thrombin: $K_m = 0.9 \cdot 10^{-5} \text{ mol/L}$
 $V = 2.2 \cdot 10^{-7} \text{ mol/min} \cdot \text{NIH-U}$

Both determined at 37°C in 2.5 mL 0.05 mol/L Tris buffer pH 8.3, I 0.15.

STANDARDIZATIONAn activity of $\Delta A/\text{min} = 0.05$ (37°C) is obtained by using 0.1 mmol/L substrate and:

- 0.03 NIH-U/mL of bovine thrombin (Roche or Parke-Davis)
- 0.04 U/mL of human thrombin (MRC standard thrombin)

Aprotinin (Trasylof®) may be added to the buffer in a concentration of 75 KIU/L in order to inhibit other activities than that of thrombin.

Note: For thrombin standardization against the MRC-standard the natural substrate fibrinogen, is recommended as the primary substrate. The clotting and amidolytic activities of degraded thrombins do not always develop in parallel.

APPLICATIONS

The substrate has been used for the determination of:

- Prothrombin in plasma (1,2)
- Antithrombin in plasma (3,4)
- Platelet factor 3 in plasma (5,6)
- Heparin in plasma (7)

**DEUTSCH****S-2238™**

Nur für Laborzwecke

S-2238 ist ein chromogenes Substrat für Thrombin.

ZUSAMMENSETZUNG

Jedes Fläschchen enthält 25 mg chromogenes Substrat S-2238 und 120 mg Mannitol als Füllstoff.

CHEMIE

Chemischer name: H-D-Phenylalanyl-L-Pipecoyl-L-Arginin-Paranitroanilind dihydrochlorid


Formel: H-D-Phe-Pip-Arg-pNA · 2HCl

Molekulargewicht: 625,6

$\epsilon_{316\text{ nm}}$: $1,27 \cdot 10^4 \text{ mol}^{-1} \cdot \text{L} \cdot \text{cm}^{-1}$

Löslichkeit: > 10 mmol/l in H₂O

Stabilität: Substanz: Bis zum, Verfalldatum haltbar. Die Substanz ist bei 2-8°C bis zum angegebenen Verfalldatum stabil. Sie darf keinem Licht ausgesetzt werden. Sie ist hygroscopisch und sollte trocken gelagert werden.

CHROMOGENIX

**Chromogenix-
Instrumentation Laboratory SpA**
V.le Monza 338 - 20128 Milano (Italy)

301929R0



REFERENCES

1. AXELSSON G et al.: Prothrombin determination by means of chromogenic peptide Substrate. *Thromb Haemost* 36, 517 (1976).
2. BERGSTRÖM K & EGBERG N: Determination of plasma prothrombin using a synthetic chromogenic substrate. Paper read at the 2nd European Congress on Clinical Chemistry, October 1976.
3. ABILGAARD et al.: Antithrombin (heparin cofactor) assay with "new" chromogenic substrates. *Thromb Res II*, 549-553 (1977).
4. Chromogenix AB: Determination of antithrombin-heparin cofactor in plasma. Laboratory Instruction.
5. SANDBERG H & ANDERSSON L O: A highly sensitive assay of platelet factor 3 using a chromogenic substrate. *Thromb Res* 14, 113-124 (1979).
6. SANDBERG H et al.: Determination of platelet factor 3 in whole blood by a chromogenic peptide substrate assay. *Thromb Res* 18, 871-882 (1980).
7. Chromogenix AB: Determination of heparin in plasma. Laboratory Instruction.